

GOVERNMENT OF KERALA  
Department of Agriculture

SHORT TENDER NOTICE

No. CC (2) BFL 1(2)/ 2012-13 (4)

Date: 05-09-2012

Sealed competitive and super-scribed Tenders are invited from competent dealers for the supply of "Moisture meter" to BioFertilizer Laboratory, Parottukonam, Thiruvananthapuram.

Tender No. with PAC	: 4 / 2012, Rs. 75,000.00
Date of commencement of sale of Tender Forms	: 18-09-2012
Closing Date of Sale of Tender Forms	: 17-10-2012, 4 P M
Last Date & Time of Receipt of Tender	: 18-10-2012, 2 P M
Date & Time of Opening of Tender	: 18-10-2012, 3.30 P M
Cost of Tender Form (Original)	: Rs 400 + VAT@ 4%- Rs.416/-
Cost of Tender Form (Duplicate)	: Rs 200 + VAT@ 4%- Rs 208/-
Date up to which the rates are to remain firm for acceptance:	: 31-03-2013

EMD: 1% of PAC in the form of Demand Draft drawn in favour of the undersigned payable at the State Bank of Travancore, Vellayambalam.

Security Deposit: 5% of the total contract amount inclusive of EMD is to be furnished as above.  
Period of Supply: With in 30 days of receipt of Supply Order.

The sealed cover containing the Tender should be super scribed as the "Tender for supply of Moisture meter for State Bio Fertilizer Laboratory, Thiruvananthapuram".

Late Tenders will not be accepted. Tenders will be opened in the presence of such Tenderers or their authorized representatives who may be present at that time.

The Price of the Tender Forms shall be received only by cash or money order (postal charges extra Rs 50/-). Tenders should be submitted in the prescribed form along with an agreement in the prescribed format in Kerala Stamp paper worth Rs 100/-.


Details with respect to the Tender notice and the materials intended to be purchased can be obtained from the office of the undersigned at working hours or from the website [www.keralaagriculture.com](http://www.keralaagriculture.com) or [www.karshikakeralam.com](http://www.karshikakeralam.com) of the Department of Agriculture.

The rules and regulations prescribed for tenders by the Govt of Kerala shall be applicable to this Tender also.

Designation and address of the Officer  
From whom the Tender Forms are to be obtained  
and to whom the Tenders are to be sent:


Chief Soil Chemist  
Central Soil & Plant Health Center  
Parottukonam, Nalanchira P O  
Thiruvanthapuram - 695015



  
Chief Soil Chemist  
CHIEF SOIL CHEMIST  
Central Soil & Plant Health Centre  
Nalanchira, P.O. Trivandrum- 15

TENDER NO. 4/2012ANNEXURE ISchedule of Materials

Item with specificatio.	Quantity
Moisture Meter 400W straight halogen lamp heating system with SRA filter and SHS weighing technology Max. Sample weight capacity - 51 gm Weight resolution - 0.001 g Moisture content display - 0.01% / 0.1% Heating technology - Halogen lamp Moisture content accuracy - 0.1% - 0.02% (over 1g & 5g respectively) Drying temperature - 50° C - 200° C Measurement mode - Moisture content( wet / dry base)/ dry content/ratio/weight Display type - Large VFD Operating temperature - 5° C - 40° C(41-104° F) less than 85% RH Sample pan size - 85mm Standard accessories - sample pans-20, pan handles-2, spoons, test sample, power cable, glass fiber sheet, fuse, dust cover, display cover etc.	1 No.



  
 Chief Soil Chemist  
 CHEMIST  
 Central Soil & Plant Health Centre  
 Nalanchira, P.O, Trivandrum- 15

കേരള സർക്കാർ  
കൃഷിവകുപ്പ്

ലഘു ദർഘാസ് പരസ്യം

നമ്പർ: സിസി (2) ബി.എഫ്.എൽ 1 (2)/ 2012 13(4)

തീയതി: 05 /09/2012

തിരുവനന്തപുരത്ത് പാറോട്ടുകോണത്ത് പ്രവർത്തിക്കുന്ന ബയോഫർട്ടിലൈസർ ലബോറട്ടറിയിലേക്ക് .....മോയ്സ്ച്ചർ മീറ്റർ...നൽകാൻ സന്നദ്ധതയുള്ള യോഗ്യരായ ഡീലർമാർ/സപ്ലയർമാരിൽ നിന്നും നിശ്ചിത ഫോറത്തിൽ തയ്യാറാക്കിയ മുദ്ര വച്ച ദർഘാസുകൾ ക്ഷണിച്ചുകൊള്ളുന്നു.

- ദർഘാസ് നമ്പറും പി എ (പി)യും : 4/2012, രൂപ. 75,000.00
- ദർഘാസ് ഫോറം വിൽപ്പന ആരംഭിക്കുന്ന തീയതി : 18 / 09 / 2012.
- ദർഘാസ് ഫോറം വിൽപ്പനയുടെ അവസാന തീയതി : 17 / 10 / 2012 4 മണി.
- ദർഘാസ് സ്വീകരിക്കുന്ന അവസാന തീയതി : 18/ 10 / 2012 ഉച്ചയ്ക്ക് 2 മണി.
- ദർഘാസ് തുറക്കുന്ന തീയതിയും സമയവും : 18/ 10 / 2012 ഉച്ചയ്ക്ക് 3.30 മണി.
- ദർഘാസ് ഫോറത്തിന്റെ വില (അസ്സൽ) : രൂപ 400 + വി.നി. 4 % = രൂപ. 416.00
- ദർഘാസ് ഫോറത്തിന്റെ വില (പകർപ്പ്) : രൂപ 200 + വി.നി. 4 % = രൂപ. 208. 00
- നിരക്കുകൾ സ്ഥിരമായി നില നിൽക്കുന്നത് ഏതു തീയതി വരെ : 31 / 3 / 2013.

നിരതദ്രവ്യം: പി എ സിയുടെ 1%.(താഴെ ഒപ്പിട്ടിരിക്കുന്ന ഉദ്യോഗസ്ഥന്റെ പേരിൽ എടുത്ത സ്റ്റേറ്റ് ബാങ്ക് ഓഫ് ട്രാവൻകൂറിന്റെ വെള്ളയമ്പലം ശാഖയിൽ മാറാവുന്ന ഡിമാന്റ് ഡ്രാഫ്റ്റായി വേണം നിരതദ്രവ്യം സമർപ്പിക്കാൻ)


സുരക്ഷാനിക്ഷേപം: പി എ സിയുടെ 5%.(നിരതദ്രവ്യമായി സമർപ്പിച്ച തുക കൂടി ഉൾപ്പെടെ) സാധനങ്ങൾ സപ്ലൈ ചെയ്യേണ്ട സമയം: സപ്ലൈ ഓർഡർ ലഭിച്ച് 30 ദിവസത്തിനകം

ദർഘാസുകൾ അടക്കം ചെയ്ത മുദ്ര വച്ച കവറിനു പുറത്ത് "ബയോഫർട്ടിലൈസർ ലബോറട്ടറിയിലേക്ക് മോയ്സ്ച്ചർ മീറ്റർ നൽകുന്നതിനുള്ള ദർഘാസ്" എന്ന് രേഖപ്പെടുത്തിയിരിക്കണം. നിശ്ചിതസമയം കഴിഞ്ഞു ലഭിക്കുന്ന ദർഘാസുകൾ സ്വീകരിക്കുന്നതല്ല. നിശ്ചയിക്കപ്പെട്ട സമയത്ത് ഹാജരുള്ള ദർഘാസുകാര്യമേയോ അവരുടെ അംഗീകൃത പ്രതിനിധികളുമേയോ സാന്നിധ്യത്തിലായിരിക്കും ദർഘാസുകൾ തുറക്കുന്നത്. ദർഘാസ് ഫോറത്തിന്റെ വില പണമായോ മണി ഓർഡർ ആയോ മാത്രമേ സ്വീകരിക്കുകയുള്ളൂ. (ഫോറം അയക്കാനുള്ള തപാൽ ചാർജ്ജ് 50 രൂപ പ്രത്യേകം അടക്കേതാണ്) നിശ്ചിത ഫോറത്തിൽ സമർപ്പിക്കുന്ന ദർഘാസുകൾക്കൊപ്പം 100 രൂപയുടെ കേരള മുദ്രപത്രത്തിൽ തയ്യാറാക്കിയ നിശ്ചിത ഫോറത്തിലുള്ള കരാർ കൂടി സമർപ്പിച്ചിരിക്കണം.

ദർഘാസിനേയോ സപ്ലൈ ചെയ്യേണ്ട സാധനങ്ങളേയോ സംബന്ധിച്ച വിശദവിവരങ്ങൾ പ്രവൃത്തിസമയത്തു ഈ ഓഫീസിൽ നേരിൽ അന്വേഷിച്ചാൽ അറിയാവുന്നതാണ്. കേരള സർക്കാരിന്റെ ദർഘാസുകളെ സംബന്ധിച്ച എല്ലാ നിയമങ്ങളും ചട്ടങ്ങളും ഈ ദർഘാസിനും ബാധകമായിരിക്കുന്നതാണ്.

ദർഘാസ് ഫോറങ്ങൾ വിൽക്കുകയും ചീഫ് സോയിൽ കെമിസ്റ്റ്,  
അവ തിരികെ സ്വീകരിക്കുകയും ചെയ്യുന്ന കേന്ദ്ര മണ്ണുപരിശോധന ശാഖ,  
ഉദ്യോഗസ്ഥന്റെ പേരും മേൽവിലാസവും: പാറോട്ടുകോണം, നാലാഞ്ചിറ. പി.ഒ.

ചീഫ് സോയിൽ കെമിസ്റ്റ്  
CHIEF SOIL CHEMIST  
Central Soil & Plant Health Centre  
Nalanchira, P.O, Trivandrum- 13



TENDER NO. 4/2012

ANNEXURE I

Schedule of Materials

Item with specification	Quantity
Moisture Meter 400W straight halogen lamp heating system with SRA filter and SHS weighing technology Max. Sample weight capacity - 51 gm Weight resolution - 0.001 g Moisture content display - 0.01% / 0.1% Heating technology - Halogen lamp Moisture content accuracy - 0.1% - 0.02% (over 1g & 5g respectively) Drying temperature - 50° C - 200° C Measurement mode - Moisture content( wet / dry base)/ dry content/ratio/weight Display type - Large VFD Operating temperature -5°C - 40° C(41-104° F) less than 85% RH Sample pan size - 85mm Standard accessories - sample pans-20, pan handles-2, spoons, test sample, power cable, glass fiber sheet, fuse, dust cover, display cover etc.	1 No.



*[Signature]*  
Chief Soil Chemist

CHIEF SOIL CHEMIST  
Gandhinagar Plant Health Centre  
Nalanchira, P.O. Trivandrum- 15