



KERALA STATE
DEPARTMENT OF AGRICULTURE

State Pesticides Testing Laboratory, Parottukonam, Nalanchira.P.O
Thiruvananthapuram. 15.

Tender No. PT (1)298/2014

Dated 24-01-2015

RE-TENDER NOTICE

Sealed superscribed Tenders are invited for the “supply of High Performance Liquid Chromatography” for the use in State Pesticide Testing Laboratory, Parottukonam, Nalanchira under the scheme on Modernization of Department Laboratories 2014-15.

Estimated cost	Rs 18,00,000/- (Rupees Eighteen Lakhs only) (Detailed Specification attached)
Cost of Original tender form	Rs 2700+5 %VAT
Cost of duplicate tender form	Rs 1400+5 %VAT
Date and time upto which tender form are sold	13/02/2015 4.00P.M
Last date and time of receipt of tender	16/02/2015 2.00P.M
Date and time of opening of tender	16/02/2015 2.30P.M
E.M.D 1% of the total cost subject to a minimum of Rs.1500/- (ies, even if the 1% of the amount of tender falls below Rs.1500/-)	Demand draft drawn in favour of Deputy Director of Agriculture Cum-Senior Chemist payable at SBT Vellayambalam Thiruvananthapuram
Amount of Security	5% of the total value of the contract rounded to nearest rupee.
Time and place of supply	within 30 days of receipt of supply order from S. P.T.Lab Parottukonam
Address to which tender to be submitted	Deputy Director of Agriculture Cum-Senior Chemist State Pesticide Testing Laboratory Parottukonam, Nalanchira PO Thiruvananthapuram-15

Tender form, special condition and further detail if any required can be had from the office of the undersigned during office hours on all working days. The price of tender forms will be received only by cash remittance or Money orders. Cheques Drafts and Postage stamps etc will not accepted

The submission of tender should be strictly in accordance with special condition and as per rules and regulation prescribed by the government



Annie George.
Deputy Director of Agriculture
Deputy Director of Agriculture
Cum Senior Chemist
Pesticide Testing Laboratory
Parottukonam, Nalanchira P.O
Thiruvananthapuram-15

**GOVERNMENT OF KERALA
DEPARTMENT OF AGRICULTURE**

State Pesticide Testing Laboratory
Parottukonam, Nalanchira PO
Thiruvananthapuram-695015
Email:-ddaptl.agri@gmailcom
Phone No-04712530867

RE-TENDER NOTICE

No.PT (1) 298/2014

Date:-24-01-2015

Sealed competitive and superscribed tenders are invited from competent dealers for the "supply of High Performance Liquid Chromatography" for State Pesticide Testing Laboratory, Parottukonam, Nalanchira, Thiruvananthapuram

Re-Tender Notification Number	PT (1) 298/2014
Purpose of Tender	Supply of High Pressure Liquid Chromatography(HPLC)
Name and address of Office	Deputy Director of Agriculture State Pesticide Testing Laboratory Parottukonam, Nalanchira P.O Thiruvananthapuram
Phone Number	0471-2530867
Cost of Tender form	Rs 2700+5%VAT
Total Cost	Rs.18,00,000/- (Rupees Eighteen Lakhs only)
Earnest money Deposit	1% of quoted cost or minimum amount Rs. 1500/- whichever is higher in the Demand Draft drawn in signed payable at SBT,
form of Favour of the under Vellayambalam	
Last date of issue of Tender form,	13-02-2015, 4.00 PM
Last date & Time of receipt of Tender	16-02-2015, 2.00 PM
Date & Time of Opening of Tender	16-02-2015, 2.30 PM



Aruna George
Deputy Director of Agriculture
Cum Senior Chemist
Pesticide Testing Laboratory
Parottukonam, Nalanchira P.O
Thiruvananthapuram-15.

Detailed Specifications for High Performance Liquid Chromatography (HPLC) With UV-Vis detector

The HPLC must be a microprocessor controlled, with extensive self diagnostics and can be operated by an external PC through chromatography software besides the following configuration and specifications.

SL No	Criteria	Details
1	The Pump	<ul style="list-style-type: none"> • with binary solvent delivery pump • Should deliver constant and pulse free solvent ranging from 0.01ml to 10.0ml/min with a precision of 0.075% RSD or better. • The effective delay volume should be less than 700 micro liters independent of back pressure. • Should meet a flow accuracy of $\pm 1\%$ or 10 ul/min, whichever is greater. • should have a pressure range of 0 to 5000 psi or more for entire flow range • the solvent delivery should provide a retention time reproducibility of less than 0.3 RSD • should have the ability to maintain the custom defined/programmed solvent com • Should have a composition accuracy of $\pm 0.5\%$ absolute, independent of backpressure. • Should have the facility to calibrate the flow and readjust the real time flow values. • Should have the mechanism for auto purging • Should have the facility to monitor the pump status like flow, pressure etc.,
2	In line degasser	An integrated degasser to de gas in line of all the four solvents be supplied which works continuously and efficiently
3	The Detector: UV-Vis. Detector	<ul style="list-style-type: none"> • Should have a wave length range of 190 – 700 nm • Should have a band width of 2 nm or better • Should have a wave length accuracy of 1 nm • Should have a very good reproducibility of absorbance at least 0.0001 or better • Should have a linearity of 5% at 1.5 AU propylparaben, at 257 nm. • Should have a base line Noise less than 5.0×10^5 AU, 230 nm. • Should have a drift less than $\pm 5 \times 10^4$ AU/hr/ ° C, 230 nm. • Should have the lamp optimization facility • Should have a taper slit flow cell. • Should have self diagnostics to check the system function and calibration • Should have an easy operation to calibrate the system • Should be able to function individual or through the chromatography software operated externally through a PC

4	Columns	1.C ₁₈ column 250mm X 4.6mm id, 10mic 1 No. 2.C ₈ Column 250mm X 4.6mm id, 5mic 1 No.
5	Guard Column	Suitable guard column sets (2 no's) with C ¹⁸ inserts should be supplied along with the instrument
6	Injector Port	Rheodyne fixed loop injector assembly with trigger type should be supplied along with the instrument. Should be provided with variable pre calibrated sample loops with maintenance kit
7	Solvent filtration kit	All glass solvent filtration apparatus with suitable filter membranes to filter aqueous and organic solvents should be supplied
8	Vacuum pump	Vacuum pump must be a dry/PTFE coated diaphragm based, operational under 220V/50Hz and constructed with chemical resistant materials. Should have an output of vacuum of 800 mm Hg , or better Should have pressure gauge/ monitor
9	Sample clarification Kit	Two sets of SS sample clarification kits along with filter membranes and other accessories should be supplied
10	Data Handling Software and PC	The personal computer with latest processor/configuration with LCD monitor and Laser jet Printer should be supplied along with the instrument. The latest software which is compatible with the chromatography software should be supplied. Original windows latest software which is compatible with the chromatography software be supplied with the system. The chromatography software should be of latest of its version and be able to collect the data from two Detectors simultaneously. The software should facilitate easy method building, data integration, Quantization in % age and custom report generation. The software should facilitate manual as well as auto integration of the data. The software should have all the features required for audit trail under GLP/GMP with a password locking facility. The software should have a real time display of the chromatogram and the process sample details.

11	Delivery and Installation	State Pesticide Testing Laboratory, Parottukonam, Nalanchira PO. Thiruvananthapuram
	A	List of the users of offered system with accessories in South India should be provided
	B	Should not be of obsolete models. Availability of spares including optic and electronic components should be guaranteed for a minimum period of ten years
	C	All supplied system should be from the same vendor and warranty applicable for all system inclusive of accessories
	D	Installation, commissioning and training on routine maintenance and operation/analytical support should be provided by factory trained application Chemist and Engineer
	E	Relevant catalogue with specifications of the offered models to be submitted along with the tender documents
	F	All necessary operation manual and maintenance manual to be provided in hard and soft copies along with the equipments
	G	A detailed list of the model quoted satisfying the requisites Tender specifications should be supplied in compliance format
	H	The rate quoted should include all expenditure including taxes, installation, Transport, loading and unloading at State Pesticide Testing Laboratory, TVPM
	I	Required demonstrations and trainings should be given free of cost in the warranty period at State Pesticide Testing Laboratory, Thiruvananthapuram
	J	AMC after warranty period for 5 years and separate comprehensive Maintenance Contract (CMC) for 5 years should be quoted



Annie George
 Deputy Director of Agriculture
 Cum Senior Chemist
 Pesticide Testing Laboratory
 Parottukonam, Nalanchira P.O.
 Thiruvananthapuram-15.

SPECIAL CONDITIONS


Part of Tender Notice No. PT(1) 298/2014, dated. 24-01-2015

(To be returned duly signed along with tender forms)

1. The materials supplied should be in strict conformity with the specifications given in the list of articles appended with tender notice.
2. Quantity shown in the tender notice is approximate. The successful tenderer is bound to supply the actual numbers shown in the supply order (which may be more or less than the published quantity) during the contract period at the accepted rates and conditions.
3. The rates quoted should be inclusive of all taxes (if applicable), Loading, unloading charges etc. for delivery at State Pesticide Testing Laboratory, and cost of installation of the equipment (without incurring any extra cost for this department.)
4. Tender should be submitted in the prescribed form along with an agreement in Kerala Stamp Paper worth Rs.100/- (One hundred only)
5. Special condition and supply order will form part of the agreement and special condition should be returned with tender form duly signed in proper places by the tenderer.
6. Before supplying commences, the successful tenderer should execute an agreement in Kerala stamp paper worth Rs.100/- (One hundred only) in prescribed form. He will also have to deposit a security of 1% of tendered amount which is inclusive of Earnest Money Deposit furnished. The security should be in the form of Demand Draft drawn in favour of Deputy Director of Agriculture cum Senior Chemist, State Pesticide Testing Lab, Parottukonam, Trivandrum Payable at State Bank of Travancore, Vellayambalam, Trivandrum.
7. The rates quoted should be firm up to 31-03-2015 and contract is valid up to 31-03-2015.
8. The supply should be completed within 30 days of receipt of supply order and claim settled at the latest by 31-03-2015.
9. Undersigned is at liberty to test and satisfy the quality and quantity of supplied articles and the contractor is liable to replace any item damaged during such testing at his cost.

10. The contractor or his agent should be present at the time of opening of tenders.
11. The contractor or his agent should be present at the time of supply and should entrust the articles to the officers concerned at this laboratory. The transportation, loading, and unloading etc. should be met by the contractor.
12. On completion of supply, the invoice in triplicate should be drawn in favour of Deputy Director of Agriculture cum Senior Chemist, State Pesticide Testing Laboratory, Thiruvananthapuram for effecting payment.
13. Separate bills should be sent for each order.
14. The payment in respect of equipments will be released only after satisfactory installation and demonstration of instruments by firm.
15. In case of any dispute, the decisions of Deputy Director of Agriculture Cum Senior Chemist will be final.
16. In case the date prescribed for opening of tender falls as a holiday it will be postponed to the very next working day without any further publicity by notices.
17. Violation of any rules / special conditions / directions of supply order by contractor will lead to rejection of tender.
18. Deputy Director of Agriculture cum Senior Chemist reserves the right to reject any tenders / cancel the purchase of any item / at his discretion without assigning any reason.
19. This order is placed on behalf of the Govt. of Kerala and the C.S.T. declaration in 'D' Form will be issued by supplied Officer, wherever necessary.

SIGNATURE OF CONTRACTOR
(Name of the Firm Seal)


DY.DIRECTOR OF AGRICULTURE
~~DEPUTY SENIOR CHEMIST~~
Cum Senior Chemist
Pesticide Testing Laboratory
Parottukonam, Nalanchira P.O.
Thiruvananthapuram-15.



കേരള സർക്കാർ

കൃഷി വകുപ്പ്

പെസ്റ്റിസൈഡ് ടെസ്റ്റിംഗ് ലബോറട്ടറി, പാറോട്ടുകോണം
തിരുവനന്തപുരം 695015

നമ്പർ: പി.റ്റി. (1) 298/2014

തീയതി 24-01-2015

റീ-ടെണ്ടർ നോട്ടീസ്

പാറോട്ടുകോണം, സ്പോർട്ട് പെസ്റ്റിസൈഡ് ടെസ്റ്റിംഗ് ലബോറട്ടറിയിലെ ഉപയോഗത്തിലേക്കായി ഹൈ പെർഫോമൻസ് ലിക്വിഡ് ക്രോമറ്റോഗ്രാഫി (എച്ച് പി എൽ സി) നൽകാൻ സന്നദ്ധതയും യോഗ്യതയും ഉള്ള അംഗീകൃത സ്ഥാപനങ്ങളിൽ നിന്നും മുദ്ര വച്ച റീ-ടെണ്ടറുകൾ ക്ഷണിച്ചുകൊള്ളുന്നു.

വിശദാംശങ്ങൾ	ദർഘാസ് നമ്പർ 1
ദർഘാസ് നമ്പർ	പി. ടി. (1) 298/2014
വാങ്ങുന്ന ഇനങ്ങൾ	ഹൈ പെർഫോമൻസ് ലിക്വിഡ് ക്രോമറ്റോഗ്രാഫി (എച്ച് പി എൽ സി) വിശദമായ സ്പെസിഫിക്കേഷൻ ഉൾപ്പെടുത്തിയിരിക്കുന്നു
അടങ്കൽ തുക	18,00,000/- ലക്ഷം രൂപ
ദർഘാസ് ഫാറത്തിന്റെ വില	2700 രൂപ + 5% വാറ്റും
ഡ്യൂപ്ലിക്കേറ്റ് ദർഘാസ് ഫാറത്തിന്റെ വില	1400 രൂപ + 5% വാറ്റും
ദർഘാസ് വിൽക്കുന്ന അവസാനതീയതി-	13/02/2015 4മണി വരെ
ദർഘാസ് സ്വീകരിക്കുന്ന അവസാന തീയതി	16/02/2015 2മണി വരെ
ദർഘാസ് തുറക്കുന്ന തീയതിയും സമയവും	16/02/2015 2.30 മണിക്ക്
നിരക്കുകൾ സ്ഥിരമായി നിലനിൽക്കുന്നത് ഏതു തീയതി വരെ	31/03/2015

നിരത ദ്രവ്യം: പി എ സിയുടെ 1%.(1500 രൂപയിൽ കുറയാതെ) (താഴെ ഒപ്പിട്ടിരിക്കുന്ന ഉദ്യോഗസ്ഥന്റെ പേരിൽ എടുത്ത, സ്പോർട്ട് ബാങ്ക് ഓഫ് ട്രാവൻകൂറിന്റെ വെള്ളയമ്പലം ശാഖയിൽ മാറാവുന്ന ഡിമാന്റ് ഡ്രാഫ്റ്റായി വേണം നിരതദ്രവ്യം സമർപ്പിക്കാൻ)

സുരക്ഷാനിക്ഷേപം: പി എ സിയുടെ 5%.(നിരതദ്രവ്യമായി സമർപ്പിച്ച തുക കൂടി ഉൾപ്പെടെ) സാധനങ്ങൾ സപ്ലൈ ചെയ്യേണ്ട സമയം: സപ്ലൈ ഓർഡർ ലഭിച്ച 30 ദിവസത്തിനകം ദർഘാസുകൾ അടക്കം ചെയ്ത മുദ്ര വച്ച കവറിനു പുറത്ത് “ഹൈ പെർഫോമൻസ് ലിക്വിഡ് ക്രോമറ്റോഗ്രാഫി (എച്ച് പി എൽ സി). നൽകുന്നതിനുള്ള ദർഘാസ്” എന്ന് രേഖപ്പെടുത്തിയിരിക്കണം.നിശ്ചിതസമയം കഴിഞ്ഞു ലഭിക്കുന്ന ദർഘാസുകൾ സ്വീകരിക്കുന്നതല്ല.നിശ്ചയിക്കപ്പെട്ട സമയത്ത് ഹാജരുള്ള ദർഘാസുകാരുടെയോ അവരുടെ അംഗീകൃത പ്രതിനിധികളുടെയോ സാന്നിധ്യത്തിലായിരിക്കും ദർഘാസുകൾ തുറക്കുന്നത്.

ദർഘാസ് ഫോറത്തിന്റെ വില പണമായോ മണി ഓർഡർ ആയോ മാത്രമേ സ്വീകരിക്കുകയുള്ളൂ. (ദർഘാസ് ഫോറം തപാലിൽ ആവശ്യമുള്ളവർ തപാൽ ചാർജ്ജ് 100 രൂപ പ്രത്യേകം അടക്കേണ്ടതാണ്) നിശ്ചിത ഫോറത്തിൽ സമർപ്പിക്കുന്ന ദർഘാസുകൾക്കൊപ്പം 100 രൂപയുടെ കേരള മുദ്രപത്രത്തിൽ തയ്യാറാക്കിയ നിശ്ചിത ഫോറത്തിലുള്ള കരാർ കൂടി സമർപ്പിച്ചിരിക്കണം.

ദർഘാസിനേയോ സപ്ലൈ ചെയ്യേണ്ട സാധനങ്ങളേയോ സംബന്ധിച്ച വിശദവിവരങ്ങൾ പ്രവൃത്തിസമയത്തു ഈ ഓഫീസിൽ നേരിൽ അന്വേഷിച്ചാൽ അറിയാവുന്നതാണ്. കേരള സർക്കാരിന്റെ ദർഘാസുകളെ സംബന്ധിച്ച എല്ലാ നിയമങ്ങളും ചട്ടങ്ങളും ഈ ദർഘാസിനും ബാധകമായിരിക്കുന്നതാണ്.

ദർഘാസിനേയോ സപ്ലൈ ചെയ്യേണ്ട സാധനങ്ങളേയോ സംബന്ധിച്ച വിശദവിവരങ്ങൾ പ്രവൃത്തിസമയത്തു ഈ ഓഫീസിൽ നേരിൽ അന്വേഷിച്ചാൽ അറിയാവുന്നതാണ്. കേരള സർക്കാരിന്റെ ദർഘാസുകളെ സംബന്ധിച്ച എല്ലാ നിയമങ്ങളും ചട്ടങ്ങളും ഈ ദർഘാസിനും ബാധകമായിരിക്കുന്നതാണ്.

ദർഘാസ് ഫോറങ്ങൾ വിൽക്കുകയും അവ തിരികെ സ്വീകരിക്കുകയും ചെയ്യുന്ന ഉദ്യോഗസ്ഥന്റെ പേരും മേൽവിലാസവും: ഡെപ്യൂട്ടി ഡയറക്ടർ ഒഫ് അഗ്രികൾച്ചർ പെസ്റ്റിസൈഡ് ടെസ്റ്റിംഗ് ലാബോറട്ടറി പാറോട്ടുകോണം, തിരുവനന്തപുരം.695 015.



ഡെപ്യൂട്ടി ഡയറക്ടർ
Deputy Director of Agriculture
Cum Senior Chemist
Pesticide Testing Laboratory
Parottukonam, Nalanchira P.O.
Thiruvananthapuram-15.

- വകുപ്പിന്റെ പേര് - കൃഷിവകുപ്പ്
- ദർഘാസിന്റെ നമ്പർ - പി. റ്റി. (1) 298/2014
- ഓഫീസിന്റെ മേൽവിലാസവും ഫോൺനമ്പറും - ഡെപ്യൂട്ടി ഡയറക്ടർ ഒഫ് അഗ്രികൾച്ചർ പെസ്റ്റിസൈഡ് ടെസ്റ്റിംഗ് ലാബോറട്ടറി പാറോട്ട്കോണം, തിരുവനന്തപുരം 0471-2530867
- ദർഘാസിന്റെ ഉദ്ദേശ്യം - ഹൈ പെർഫോമൻസ് ലിമിറ്റഡ് ക്രോമറ്റോഗ്രാഫി (എച്ച് പി എൽ സി). സപ്ലൈ ചെയ്യാൻ
- ദർഘാസ് ഫാറത്തിന്റെ വില - 2700 രൂപ + വാറ്റ് 5%
- അടക്കൽ തുക - 18,00,000/ ലക്ഷം രൂപ
- നിരത ദ്രവ്യം - അടങ്കലിന്റെ 1%
- ദർഘാസ് വിൽക്കുന്ന അവസാനതീയതി- - 13/02/2015 4 മണിവരെ



Annie George
 Deputy Director of Agriculture
 Cum Senior Chemist
 Pesticide Testing Laboratory
 Parottukonam, Nalanchira P.O
 Thiruvananthapuram-15.